

SUPERVISOR TEST

ORANGETOWN TOWN
POLICE DEPT.
DATAMASTER dmt: 115206

Date: 11/24/2022

OPERATOR NAME:
JOSEPH W SULLIVAN
Permit #: 46948
Operator Agency/Dept: ORANGETOWN TOWN P.D.
Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:09
INTERNAL STANDARD	VERIFIED	16:09
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	16:09
BLANK TEST	0.00	16:10
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	16:10
BLANK TEST	0.00	16:11
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	16:11
BLANK TEST	0.00	16:12

Making a false statement which a person does not believe to be true is a Class A misdemeanor in violation of Section 210.45 of the penal law of the State of New York.

Date 11/24/2022
Signature [Handwritten Signature]

