

**SUPERVISOR TEST**

-----  
ORANGETOWN TOWN  
POLICE DEPT.  
DATAMASTER dmt: 115206  
-----

Date: 11/24/2023

OPERATOR NAME:  
A PALAZOLO  
Permit #: 42477  
Operator Agency/Dept: ORANGETOWN TOWN P.D.  
Reference Standard #: 23270

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:42
INTERNAL STANDARD	VERIFIED	09:42
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	09:43
BLANK TEST	0.00	09:44
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	09:44
BLANK TEST	0.00	09:45
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	09:45
BLANK TEST	0.00	09:46

**Making a false statement which a person does not believe to be true is a Class A misdemeanor in violation of Section 210.45 of the penal law of the State of New York.**

Date 11/24/23

Signature *D/S* #321

