

SOLUTION CHANGE

ORANGETOWN TOWN
POLICE DEPT.
DATAMASTER dmt: 115206

Date: 10/01/2021

OPERATOR NAME:
JOSEPH W SULLIVAN
Permit #: 46948
Operator Agency/Dept: ORANGETOWN TOWN P.D.
Reference Standard #: 21280

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:07
INTERNAL STANDARD	VERIFIED	16:07
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	16:07
BLANK TEST	0.00	16:08
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	16:08
BLANK TEST	0.00	16:09
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	16:09
BLANK TEST	0.00	16:10

