

SUPERVISOR TEST

ORANGETOWN TOWN
POLICE DEPT.
DATAMASTER dmt: 115206

Date: 10/02/2024

OPERATOR NAME:
T AMBROSE
Permit #: 53435
Operator Agency/Dept: ORANGETOWN TOWN P.D.
Reference Standard #: 24190

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:42
INTERNAL STANDARD	VERIFIED	10:42
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	10:43
BLANK TEST	0.00	10:44
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	10:44
BLANK TEST	0.00	10:45
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	10:45
BLANK TEST	0.00	10:46

Making a false statement which a person does not believe to be true is a Class A misdemeanor in violation of Section 210.45 of the penal law of the State of New York.

10/3/24

Signature *T. Ambrose*

