

SOLUTION CHANGE

ORANGETOWN TOWN
POLICE DEPT.
DATAMASTER dmt: 115206

Date: 10/03/2022

OPERATOR NAME:
JOSEPH W SULLIVAN
Permit #: 46948
Operator Agency/Dept: ORANGETOWN TOWN P.D.
Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:15
INTERNAL STANDARD	VERIFIED	08:15
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:15
BLANK TEST	0.00	08:16
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:16
BLANK TEST	0.00	08:17
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:17
BLANK TEST	0.00	08:18

