

SUPERVISOR TEST

ORANGETOWN TOWN
POLICE DEPT.
DATAMASTER dmt: 115206

Date: 10/20/2023

OPERATOR NAME:
A PALAZOLO
Permit #: 42477
Operator Agency/Dept: ORANGETOWN TOWN P.D.
Reference Standard #: 23270

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:01
INTERNAL STANDARD	VERIFIED	15:01
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	15:02
BLANK TEST	0.00	15:03
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	15:03
BLANK TEST	0.00	15:04
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	15:04
BLANK TEST	0.00	15:05

Making a false statement which a person does not believe to be true is a Class A misdemeanor in violation of Section 210.45 of the penal law of the State of New York.

Date 10/20/23

Signature *[Signature]* #224

