

SUPERVISOR TEST

ORANGETOWN TOWN
POLICE DEPT.
DATAMASTER dmt: 115206

Date: 10/26/2022

OPERATOR NAME:
JOSEPH W SULLIVAN
Permit #: 46948
Operator Agency/Dept: ORANGETOWN TOWN P.D.
Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:55
INTERNAL STANDARD	VERIFIED	15:55
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	15:55
BLANK TEST	0.00	15:56
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	15:56
BLANK TEST	0.00	15:57
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	15:58
BLANK TEST	0.00	15:58

Making a false statement which a person does not believe to be true is a Class A misdemeanor in violation of Section 210.45 of the penal law of the State of New York.

Date 10/26/2022
Signature [Handwritten Signature]

