

SUPERVISOR TEST

ORANGETOWN TOWN  
POLICE DEPT.  
DATAMASTER dmt: 115206

Date: 09/13/2024

OPERATOR NAME:  
T AMBROSE  
Permit #: 53435  
Operator Agency/Dept: ORANGETOWN TOWN P.D.  
Reference Standard #: 24190

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:10
INTERNAL STANDARD	VERIFIED	19:10
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	19:10
BLANK TEST	0.00	19:11
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	19:12
BLANK TEST	0.00	19:12
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	19:13
BLANK TEST	0.00	19:14

Making a false statement which a person does not believe to be true is a Class A misdemeanor in violation of Section 210.45 of the penal law of the State of New York.

Date 9/13/24  
Signature T Ambrose 258

