

SUPERVISOR TEST

ORANGETOWN TOWN
POLICE DEPT.
DATAMASTER dmt: 115206

Date: 09/15/2021

OPERATOR NAME:
JOSEPH W SULLIVAN
Permit #: 46948
Operator Agency/Dept: ORANGETOWN TOWN P.D.
Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:05
INTERNAL STANDARD	VERIFIED	16:05
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	16:05
BLANK TEST	0.00	16:06
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	16:06
BLANK TEST	0.00	16:07
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	16:08
BLANK TEST	0.00	16:08

