

SUPERVISOR TEST

ORANGETOWN TOWN
POLICE DEPT.
DATAMASTER dmt: 115206

Date: 09/23/2021

OPERATOR NAME:
JOSEPH W SULLIVAN
Permit #: 46948
Operator Agency/Dept: ORANGETOWN TOWN P.D.
Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:51
INTERNAL STANDARD	VERIFIED	07:51
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	07:51
BLANK TEST	0.00	07:52
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	07:52
BLANK TEST	0.00	07:53
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	07:53
BLANK TEST	0.00	07:54

