

SUPERVISOR TEST

ORANGETOWN TOWN
POLICE DEPT.
DATAMASTER dmt: 115206

Date: 09/28/2023

OPERATOR NAME:
JOSEPH W SULLIVAN
Permit #: 46948
Operator Agency/Dept: ORANGETOWN TOWN P.D.
Reference Standard #: 23190

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:24
INTERNAL STANDARD	VERIFIED	09:24
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	09:25
BLANK TEST	0.00	09:26
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	09:26
BLANK TEST	0.00	09:27
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	09:27
BLANK TEST	0.00	09:28

Making a false statement which a person does not believe to be true is a Class A misdemeanor in violation of Section 210.45 of the penal law of the State of New York.
Date 09/28/23
Signature Sgt. J. Sullivan

