

SOLUTION CHANGE

ORANGETOWN TOWN
POLICE DEPT.
DATAMASTER dmt: 115206

Date: 04/01/2020

OPERATOR NAME:
JOSEPH W SULLIVAN
Permit #: 46948
Operator Agency/Dept: ORANGETOWN TOWN P.D.
Reference Standard #: 20040

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:57
INTERNAL STANDARD	VERIFIED	23:57
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	23:57
BLANK TEST	0.00	23:58
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	23:59
BLANK TEST	0.00	00:00
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	00:00
BLANK TEST	0.00	00:01

