

SUPERVISOR TEST

ORANGETOWN TOWN
POLICE DEPT.
DATAMASTER dmt: 115206

Date: 04/22/2020

OPERATOR NAME:
JOSEPH W SULLIVAN
Permit #: 46948
Operator Agency/Dept: ORANGETOWN TOWN P.D.
Reference Standard #: 20040

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:48
INTERNAL STANDARD	VERIFIED	23:48
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	23:48
BLANK TEST	0.00	23:49
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	23:50
BLANK TEST	0.00	23:51
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	23:51
BLANK TEST	0.00	23:52

