

SOLUTION CHANGE

ORANGETOWN TOWN  
POLICE DEPT.  
DATAMASTER dmt: 115206

Date: 08/14/2020

OPERATOR NAME:  
JOSEPH W SULLIVAN  
Permit #: 46948  
Operator Agency/Dept: ORANGETOWN TOWN P.D.  
Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:55
INTERNAL STANDARD	VERIFIED	07:56
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:56
BLANK TEST	0.00	07:57
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	07:57
BLANK TEST	0.00	07:58
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	07:58
BLANK TEST	0.00	07:59

