

SOLUTION CHANGE

ORANGETOWN TOWN  
POLICE DEPT.  
DATAMASTER dmt: 115206

Date: 12/04/2020

OPERATOR NAME:  
JOSEPH W SULLIVAN  
Permit #: 46948  
Operator Agency/Dept: ORANGETOWN TOWN P.D.  
Reference Standard #: 20430

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:39
INTERNAL STANDARD	VERIFIED	17:39
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	17:39
BLANK TEST	0.00	17:40
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	17:40
BLANK TEST	0.00	17:41
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	17:42
BLANK TEST	0.00	17:43

