

SUPERVISOR TEST

ORANGETOWN TOWN
 POLICE DEPT.
 DATAMASTER dmt: 115206

Date: 12/16/2020

OPERATOR NAME:
 JOSEPH W SULLIVAN
 Permit #: 46948
 Operator Agency/Dept: ORANGETOWN TOWN P.D.
 Reference Standard #: 20430

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:35
INTERNAL STANDARD	VERIFIED	08:35
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	08:35
BLANK TEST	0.00	08:36
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:36
BLANK TEST	0.00	08:37
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:38
BLANK TEST	0.00	08:39

