

SUPERVISOR TEST

ORANGETOWN TOWN
POLICE DEPT.
DATAMASTER dmt: 115206

Date: 02/08/2020

OPERATOR NAME:
JOSEPH W SULLIVAN
Permit #: 46948
Operator Agency/Dept: ORANGETOWN TOWN P.D.
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:28
INTERNAL STANDARD	VERIFIED	19:28
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	19:29
BLANK TEST	0.00	19:30
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	19:30
BLANK TEST	0.00	19:31
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	19:31
BLANK TEST	0.00	19:32

