

SUPERVISOR TEST

ORANGETOWN TOWN  
POLICE DEPT.  
DATAMASTER dmt: 115206

Date: 01/06/2021

OPERATOR NAME:  
JOSEPH W SULLIVAN  
Permit #: 46948  
Operator Agency/Dept: ORANGETOWN TOWN P.D.  
Reference Standard #: 20430

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:02
INTERNAL STANDARD	VERIFIED	15:02
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	15:03
BLANK TEST	0.00	15:04
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	15:04
BLANK TEST	0.00	15:05
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	15:05
BLANK TEST	0.00	15:06

