

SUPERVISOR TEST

ORANGETOWN TOWN  
POLICE DEPT.  
DATAMASTER dmt: 115206

Date: 01/15/2021

OPERATOR NAME:  
JOSEPH W SULLIVAN  
Permit #: 46948  
Operator Agency/Dept: ORANGETOWN TOWN P.D.  
Reference Standard #: 20430

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:59
INTERNAL STANDARD	VERIFIED	07:59
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	07:59
BLANK TEST	0.00	08:00
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	08:00
BLANK TEST	0.00	08:01
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	08:02
BLANK TEST	0.00	08:02

