

SUPERVISOR TEST

ORANGETOWN TOWN
POLICE DEPT.
DATAMASTER dmt: 115206

Date: 07/18/2020

OPERATOR NAME:
JOSEPH W SULLIVAN
Permit #: 46948
Operator Agency/Dept: ORANGETOWN TOWN P.D.
Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:55
INTERNAL STANDARD	VERIFIED	13:55
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	13:56
BLANK TEST	0.00	13:57
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	13:57
BLANK TEST	0.00	13:58
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	13:58
BLANK TEST	0.00	13:59

