

SUPERVISOR TEST

ORANGETOWN TOWN
POLICE DEPT.
DATAMASTER dmt: 115206

Date: 03/02/2020

OPERATOR NAME:
JOSEPH W SULLIVAN
Permit #: 46948
Operator Agency/Dept: ORANGETOWN TOWN P.D.
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:35
INTERNAL STANDARD	VERIFIED	03:36
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:36
BLANK TEST	0.00	03:37
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:37
BLANK TEST	0.00	03:38
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:38
BLANK TEST	0.00	03:39

