

SOLUTION CHANGE

ORANGETOWN TOWN  
POLICE DEPT.  
DATAMASTER dmt: 115206

Date: 03/02/2020

OPERATOR NAME:  
JOSEPH W SULLIVAN  
Permit #: 46948  
Operator Agency/Dept: ORANGETOWN TOWN P.D.  
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:57
INTERNAL STANDARD	VERIFIED	03:57
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	03:58
BLANK TEST	0.00	03:59
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	03:59
BLANK TEST	0.00	04:00
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	04:00
BLANK TEST	0.00	04:01

