

SUPERVISOR TEST

ORANGETOWN TOWN  
POLICE DEPT.  
DATAMASTER dmt: 115206

Date: 03/11/2020

OPERATOR NAME:  
JOSEPH W SULLIVAN  
Permit #: 46948  
Operator Agency/Dept: ORANGETOWN TOWN P.D.  
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:47
INTERNAL STANDARD	VERIFIED	00:47
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	00:47
BLANK TEST	0.00	00:48
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	00:48
BLANK TEST	0.00	00:49
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	00:49
BLANK TEST	0.00	00:50

