

SUPERVISOR TEST

ORANGETOWN TOWN  
POLICE DEPT.  
DATAMASTER dmt: 115206

Date: 05/10/2020

OPERATOR NAME:  
JOSEPH W SULLIVAN  
Permit #: 46948  
Operator Agency/Dept: ORANGETOWN TOWN P.D.  
Reference Standard #: 20040

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:18
INTERNAL STANDARD	VERIFIED	00:18
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	00:18
BLANK TEST	0.00	00:19
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	00:19
BLANK TEST	0.00	00:20
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	00:21
BLANK TEST	0.00	00:22

