

SOLUTION CHANGE

ORANGETOWN TOWN
POLICE DEPT.
DATAMASTER dmt: 115206

Date: 11/04/2020

OPERATOR NAME:
JOSEPH W SULLIVAN
Permit #: 46948
Operator Agency/Dept: ORANGETOWN TOWN P.D.
Reference Standard #: 20340

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:00
INTERNAL STANDARD	VERIFIED	15:00
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	15:01
BLANK TEST	0.00	15:01
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	15:02
BLANK TEST	0.00	15:03
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:03
BLANK TEST	0.00	15:04

