

SOLUTION CHANGE

ORANGETOWN TOWN
POLICE DEPT.
DATAMASTER dmt: 115206

Date: 09/13/2020

OPERATOR NAME:
JOSEPH W SULLIVAN
Permit #: 46948
Operator Agency/Dept: ORANGETOWN TOWN P.D.
Reference Standard #: 20240

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:21
INTERNAL STANDARD	VERIFIED	08:21
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:22
BLANK TEST	0.00	08:23
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:23
BLANK TEST	0.00	08:24
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:24
BLANK TEST	0.00	08:25

