

SUPERVISOR TEST

ORANGETOWN TOWN
POLICE DEPT.
DATAMASTER dmt: 115206

Date: 09/21/2020

OPERATOR NAME:
JOSEPH W SULLIVAN
Permit #: 46948
Operator Agency/Dept: ORANGETOWN TOWN P.D.
Reference Standard #: 20240

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:58
INTERNAL STANDARD	VERIFIED	16:58
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	16:58
BLANK TEST	0.00	16:59
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	16:59
BLANK TEST	0.00	17:00
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	17:01
BLANK TEST	0.00	17:01

